

A CT Quality Control Audit Program Based on Automated Analysis and Alerts

Emi Eastman; Di Zhang, Ph.D., DABR; Yifang Zhou, Ph.D, DABR

Imaging Department
Cedars-Sinai Medical Center



Purpose

- A continuous quality control (QC) program is an essential part of computed tomography (CT) quality assurance to ensure patient safety and consistent clinical performance.
- The American College of Radiology (ACR) requires that the continuous QC program be monitored by a Qualified Medical Physicist (QMP).
- In a clinical setting with multiple sites and scanners, this monitoring activity often entails a constant, detailed, and laborious review of records by a medical physicist.
- The purpose of this study was to minimize the amount of human review time by developing a self-regulating program to monitor CT QC results that can detect and report both missing and out-of-tolerance results.

Methods

1. We created a set of electronic forms using Microsoft Excel for technologists to use to record the QC results.

- Stored on a cloud drive that both technologists and physicists have access to.
- Incorporate limits specified by both the vendors and the ACR for each of the QC tests.
- The daily test includes CT number, noise, the presence of artifacts, and uniformity.
- The weekly test includes SMPTE pattern check.
- The monthly test includes visual inspections of the console and gantry, laser alignment, and vendor-recommended quantitative tests such as spatial resolution, contrast scale, slice thickness, and low contrast resolution.

GE Revolution Quality Assurance Record - Technologist Daily, Weekly and Monthly Checks																	
DATE	UNIT STATUS (select)	ARTIFACTS			DAILY CT # CHECK					WEEKLY SMPTE CHECK				TECH INITIALS			
		Free of Artifacts	PASS / FAIL	#OK (process)	Scan Type	CTF Acc. Max Abs	Noise	Top CTF	Right CTF	Uniformity Max Abs	Water CT # Uniformity	LCD 5mm	PASS / FAIL		5% Visible	95% Visible	No Aliasing
2/1/2021	OPEN	YES	PASS	OK	Axial	-1.33	4.87		2.31		2.29	3.02	PASS				
2/2/2021	OPEN	YES	PASS	OK	Helical	-1.22	4.84		2.28		2.29	3.01	PASS				
2/3/2021	OPEN	YES	PASS	OK	Axial	-1.52	4.81		3.07		2.41	3.01	PASS	YES	YES	YES	PASS
2/4/2021	OPEN	YES	PASS	OK	Helical	-1.43	4.82		2.28		2.41	3.01	PASS				
2/5/2021	OPEN	YES	PASS	OK	Axial	-0.95	4.83		2.35		2.63	3.09	PASS				
2/6/2021	OPEN	YES	PASS	OK	Helical	-1.54	4.85		2.33		2.63	3.09	PASS				
2/7/2021	OPEN	YES	PASS	OK	Axial	-1.47	4.89		2.50		2.63	3.05	PASS				
2/8/2021	OPEN	YES	PASS	OK	Helical	-1.34	4.85		2.77		2.57	3.04	PASS				
2/9/2021	OPEN	YES	PASS	OK	Axial	-1.59	4.82		2.77		2.57	3.04	PASS				
2/10/2021	OPEN	YES	PASS	OK	Axial	-1.22	4.82		2.34		2.57	3.06	PASS	YES	PASS	YES	PASS
2/11/2021	OPEN	YES	PASS	OK	Helical	-1.38	4.83		2.33		2.57	3.03	PASS				
2/12/2021	OPEN	YES	PASS	OK	Axial	-1.73	4.83		2.85		2.52	3.02	PASS				
2/13/2021	OPEN	YES	PASS	OK	Helical	-1.48	4.85		2.33		2.52	3.03	PASS				
2/14/2021	OPEN	YES	PASS	OK	Axial	-1.04	4.87		2.57		2.52	3.06	PASS				
2/15/2021	OPEN	YES	PASS	OK	Helical	-0.95	4.91		2.54		2.04	3.07	PASS				
2/16/2021	OPEN	YES	PASS	OK	Axial	-1.73	4.87		2.80		2.38	3.07	PASS				
2/17/2021	OPEN	YES	PASS	OK	Helical	-1.33	4.84		2.33		2.38	3.08	PASS	YES	YES	YES	PASS
2/18/2021	OPEN	YES	PASS	NO	Axial	-1.10	4.87		2.39		2.38	3.05	PASS				
2/19/2021	OPEN	YES	PASS	OK	Helical	-1.54	4.86		2.77		2.77	3.03	PASS				
2/20/2021	OPEN	YES	PASS	OK	Axial	-1.61	4.85		2.59		2.83	3.04	PASS				
2/21/2021	OPEN	YES	PASS	OK	Helical	-1.74	4.82		2.47		2.83	3.08	PASS				
2/22/2021	OPEN	YES	PASS	OK	Axial	-1.55	4.86		2.47		2.26	3.06	PASS				
2/23/2021	OPEN	YES	PASS	OK	Helical	-1.13	4.88		2.32		2.32	3.07	PASS	YES	YES	YES	PASS
2/24/2021	OPEN	YES	PASS	OK	Axial	-1.71	4.87		2.81		2.57	3.08	PASS				
2/25/2021	OPEN	YES	PASS	OK	Helical	-1.53	4.83		2.63		2.57	3.08	PASS				
2/26/2021	OPEN	YES	PASS	OK	Axial	-1.39	4.85		2.63		2.57	3.07	PASS				
2/27/2021	OPEN	YES	PASS	OK	Helical	-1.19	4.88		2.33		2.26	3.08	PASS				
2/28/2021	OPEN	YES	PASS	OK	Axial	-1.23	4.85		2.49		2.26	3.05	PASS				

GE Revolution Quality Assurance Record - Monthly Checks					Graphs	
DATE	2/14/2021					
Quantitative Tests	Monthly Quantitative Checks		PASS / FAIL	VALUE	TECH INITIALS	Hidden for presentation
	TEST					
	Average Contrast Scale	PASS	121.63			
	Spatial Resolution (MTF 50)	PASS	9.11			
	Spatial Resolution (MTF 10)	PASS	5.66			
	Slice Thickness (0.625 mm)	PASS	0.67			
Laser Alignment	TEST		PASS / FAIL	TECH INITIALS	Hidden for presentation	
	TEST					
Gantry Test	TEST		PASS / FAIL	TECH INITIALS	Hidden for presentation	
	TEST					
	Table Height Indicator works	PASS				
	Table position indicator works	PASS				
Console Test	TEST		PASS / FAIL	TECH INITIALS	Hidden for presentation	
	TEST					
	Exposure switch works	PASS				
	Panel switches/lights/indicators work	PASS				
Paperwork	TEST		PASS / FAIL	TECH INITIALS	Hidden for presentation	
	TEST					
	Warning labels present	PASS				
	PM Service records in Log book	PASS				

Methods

2. These digital forms are reviewed automatically by a Python-based program at a user-defined frequency.
 - Our institution uses the program:
 - **Daily** to search the forms for out of tolerance results—either a warning or failure depending on the specified upper and lower limits
 - **Weekly** to identify missing entries
 - The program also maintains a record of all out-of-tolerance results detected throughout the month in a separate file.
 - The Pandas library was used for all data manipulation and Pywin32 was used to create copies of the original files for our records

Methods

- When an out-of-tolerance result is detected, the program sends an email alert to a group of defined users.
 - The Pywin32 library was used to send out email alerts.
 - The program automatically runs at our defined frequency using Windows Task Scheduler on a desktop computer.

Daily CT QC Alert: 2021-02-24

New Fails/Warnings

Date	Machine	Test	Result
2021-02-23	AHSP Siemens CT	Water CT # Uniformity	WARNING

Current Blanks

Date	Machine	Test	Result
2021-02-16	Angeles Biograph	Daily CT # Check	Blank
2021-02-16	Angeles Biograph	Free of Artifacts	Blank
2021-02-14	CCC GE VCT	Weekly SMPTE	Blank
2021-02-19	ED Toshiba Genesis	Daily CT # Check	Blank
2021-02-19	ED Toshiba Genesis	Free of Artifacts	Blank

The email alert sent when the program detects an out-of-tolerance or blank result.

Results

The accuracy of the program was evaluated for 16 weeks, and the results were compared to human performance of the same task. In addition, the total number of missing records were compared before and after the implementation of the automated QC program to evaluate the behavior change of the technologists performing the QC.

- The results of a manual review of the QC records agreed with the results reported in the email alerts with the accuracy of 100%.
- For our enterprise health system with 16 CT scanners, implementing the automated program saves 60 minutes of a medical physicist's time per week, or 50 hours per year.
- The program detected missing entries that the human review did not capture.

Results

- After the first two weeks of implementation, the number of missing QC results detected each week decreased from 12 and 14 in the first and second week, respectively, to a median of 2 each week for the following 14 weeks.
- The continuous and detailed monitoring by the program in combination with the routine, weekly follow-up by the physicists led to more timely QC entries by the technologists and an overall decline in missing QC records.




Figure C. The number of missing QC records reported each week.

Conclusions

We successfully implemented an automated method for CT QC documentation and review. This method not only increased the productivity of medical physicists by saving a significant amount of time and avoided human errors, but also drove behavioral changes of technologists which led to significantly reduced missing records.